

logical controls, as supported by most professional societies around the world. One has to question, therefore, why Pérez-Padilla et al chose not to include any physiological controls in their study.

It is also noticeable that Pérez-Padilla et al chose only to use the American Thoracic Society's 1994 calibration protocol,<sup>2</sup> which was verification at a single flow rate, rather than the multi-point method recommended in the 2005 spirometry standard from the American Thoracic Society and European Respiratory Society.<sup>3</sup> Did Pérez-Padilla et al know that the EasyOne spirometer does not meet the requirement of the multi-point method in the 2005 standard? Why did they not mention this in the paper?

Unfortunately, the Pérez-Padilla et al paper shows a particular brand of spirometer in a good light. It is not my intention to comment on whether the device is a good one or a poor one, but instead simply to point out that publishing articles with flawed methods and incorrect conclusions can put a device in either a positive or negative light, which could have an important effect on the sales of that manufacturer. To give a misleading impression reflects badly upon what has up until now, in my opinion, been a revered journal.

#### Alan J Moore

Respiratory Physiology Service  
Department of Thoracic Medicine  
City Hospital  
Birmingham, United Kingdom

The author has current or has previously had consultancy agreements, or is receiving or has previously received educational sponsorship and/or hospitality, from the following manufacturers or vendors of spirometers or equipment related to pulmonary function testing: Beaver Medical PLC—UK, Clement Clarke International, Custo Med—Germany, Ferraris Cardiorespiratory, Medical International Research (MIR)—Rome, ndd Medical Technologies, and Viasys Healthcare.

#### REFERENCES

1. Pérez-Padilla R, Vázquez-García JC, Márquez MN, Jardim JR, Pertuzé J, Lisboa C, et al. The long-term stability of portable

spirometers used in a multinational study of the prevalence of chronic obstructive pulmonary disease. *Respir Care* 2006;51(10):1167–1171.

2. American Thoracic Society. Standardization of spirometry, 1994 update. *Am J Respir Crit Care Med* 1995;152(3):1107–1136.
3. Miller MR, Hankinson J, Brusasco V, Burgos F, Casaburi R, Coates A, et al. Standardisation of spirometry. *Eur Respir J* 2005;26(2):319–338.

#### *The authors respond:*

Mr Moore is correct that we did not test daily linearity, as recommended in the 2005 American Thoracic Society/European Respiratory Society (ATS-ERS) standard<sup>1</sup> (3 syringe injections, with different syringe-emptying speeds). As we described in our paper,<sup>2</sup> we performed one 3-L daily calibration check during the survey, which took place during 2003 and 2004. The available ATS standards were from 1994.<sup>3</sup> However, at the end of the survey, we observed that 3-L calibration checks were not done at one single flow, (as Mr Moore suggested), but with a wide variety of syringe flows (see Fig. 2 in our paper<sup>2</sup>). The peak flows we obtained during the syringe injections in 5% of the calibration checks were higher than 8 L/s (maximum 9.8 L/s) and in 5% of the calibration checks were lower than 2.2 L/s (minimum 0.4 L/s). But that wide range of flows did not significantly influence the measured volume, as can be seen in our Figure 2 and as described in the text. That the flow did not significantly affect the volume requires linearity, at least in that range of calibration flows, and is a piece of information available in our survey and maybe in others, with one syringe injection per day. Of course, the spirometers were not tested with all the flows shown in our Figure 2, but each spirometer had a variety of injection flows (mean flow range  $5.9 \pm 2.0$  L/s), applied on different days during the survey, and had correct volumes. This is the sense in which we used the term “linearity,” and not that described in the 2005 ATS-ERS standards.<sup>3</sup> Re-

cently, Walters et al published results similar to ours.<sup>4</sup>

We communicated our experience with 70 EasyOne spirometers during a survey done house-by-house, following a strict quality-control protocol. We have no doubt that other handheld spirometers can have at least a similar performance to the EasyOne spirometers we reported on. Researchers experienced in the long-term use of other devices should publish their results to provide potential users with this valuable information. Long-term calibration stability, the main issue in our paper, is a necessity if we want to expand the use of spirometry, especially in general practice.

Recently we collected 47 of the EasyOne spirometers used in the survey (in Mexico City, Montevideo, Sao Paulo, and Santiago) and tested their flow linearity with a flow-volume calibrator (FVC 3000, Jones Medical Instrument, Oak Brook, Illinois), with 17 flow points, ranging from  $< 1$  L/s to 16 L/s. The remaining 10 spirometers were not tested: 3 were out of order and 7 were unavailable. This calibration was after 2–3 years of use. The overall concordance correlation coefficient<sup>5</sup> between the syringe and the calibrator-measured flow was 0.995 (95% confidence interval 0.994–0.996) and the 95% limits of agreement<sup>6</sup> were between  $-0.431$  and  $0.663$  L/s. The calibration of 13 additional spirometers used in Caracas was adequate when tested with a 3-L syringe and 3 different flows, as required by current standards.

Finally, as we stated in the paper, none of the authors has a commercial relationship with the manufacturer of the EasyOne spirometer, so none of us will benefit if EasyOne sales increase.

#### Rogelio Pérez-Padilla MD

on behalf of the  
PLATINO Study Group  
Instituto Nacional de Enfermedades  
Respiratorias  
Distrito Federale  
Mexico City, Mexico

The authors report no conflict of interest related to the content of this letter.

**REFERENCES**

1. Miller MR, Hankinson J, Brusasco V, Burgos F, Casaburi R, Coates A, et al. Standardisation of spirometry. *Eur Respir J* 2005;26(2):319–338.
2. Pérez-Padilla R, Vázquez-García JC, Márquez MN, Jardim JR, Pertuzé J, Lisboa C, et al. The long-term stability of portable spirometers used in a multinational study of the prevalence of chronic obstructive pulmonary disease. *Respir Care* 2006;51(10):1167–1171.
3. American Thoracic Society. Standardization of spirometry, 1994 update. *Am J Respir Crit Care Med* 1995;152(3):1107–1136.
4. Walters JA, Wood-Baker R, Walls J, Johns DP. Stability of the EasyOne ultrasonic spirometer for use in general practice. *Respirology* 2006;11:306–310.
5. Lin LI. A concordance correlation coefficient to evaluate reproducibility. *Biometrics* 1989;45(1):255–268.
6. Bland JM, Altman DG. Statistical methods for assessing agreement between 2 methods of clinical measurement. *Lancet* 1986;1(8476):307–310.



Ohio 2200 Pulmonary Function Testing System  
Advertisement by Ohio Medical Products, Madison, Wisconsin  
*RESPIRATORY CARE* Vol 16, No 6, November-December 1971