

AARC Clinical Practice Guideline

Capnography/Capnometry during Mechanical Ventilation— 2003 Revision & Update

CO₂ MV 1.0 PROCEDURE:

Capnography comprises the continuous analysis and recording of carbon dioxide concentrations [CO₂] in respiratory gases. Although the terms capnography and capnometry are sometimes considered synonymous, capnometry suggests measurement (ie, analysis alone) without a continuous written record or waveform. Capnographic waveforms may be time-based or volume-based.

CO₂ MV 2.0 DESCRIPTION/DEFINITION:

For the purposes of this Guideline, capnography refers to the evaluation of the [CO₂] in the respiratory gases of mechanically ventilated patients. A capnographic device incorporates one of two types of analyzers: mainstream or sidestream.^{1,2} Mainstream analyzers insert a sampling window into the ventilator circuit for measurement of CO₂, whereas a sidestream analyzer aspirates gas from the ventilator circuit, and the analysis occurs away from the ventilator circuit. Analyzers utilize infrared, mass or Raman spectra, or a photoacoustic spectra technology.³⁻⁵ Flow measuring devices are utilized in volume-based capnographs.

CO₂ MV 3.0 SETTING:

This procedure may be performed by trained health care personnel in any setting in which mechanically ventilated patients are found—for example, the hospital, the extended care facility, or during transport.

CO₂ MV 4.0 INDICATIONS:

On the basis of available evidence, capnography should not be mandated for all patients receiving mechanical ventilatory support, but it may be indicated for:

4.1 Evaluation of the exhaled [CO₂], especially end-tidal CO₂, which is the maximum partial pressure of CO₂ exhaled during a tidal breath

(just prior to the beginning of inspiration) and is designated P_{etCO₂}.^{6,7}

4.2 Monitoring severity of pulmonary disease⁵ and evaluating response to therapy, especially therapy intended to improve the ratio of dead space to tidal volume (\dot{V}_D/\dot{V}_T)^{8,9} and the matching of ventilation to perfusion (\dot{V}/\dot{Q}),¹⁰ and, possibly, to increase coronary blood flow^{6,11,12}

4.3 Use as an adjunct to determine that tracheal rather than esophageal intubation has taken place (Low or absent cardiac output may negate its use for this indication.);¹³⁻¹⁸ colorimetric CO₂ detectors are adequate devices for this purpose.¹⁹

4.4 Continued monitoring of the integrity of the ventilatory circuit, including the artificial airway^{13,16}

4.5 Evaluation of the efficiency of mechanical ventilatory support by determination of the difference between the arterial partial pressure for CO₂ (P_{aCO₂}) and the P_{etCO₂}.^{20,21}

4.6. Monitoring adequacy of pulmonary, systemic, and coronary blood flow^{11,12,22-26}

4.6.1 Estimation of effective (nonshunted) pulmonary capillary blood flow by a partial rebreathing method²⁷⁻²⁹

4.6.2 Use as an adjunctive tool to screen for pulmonary embolism (Evidence for the utility of dead space determinations as a screening tool for pulmonary embolism is at present not conclusive.)^{30,31}

4.6.3 Monitoring the matching of ventilation to perfusion during independent lung ventilation for unilateral pulmonary contusion³²

4.7 Monitoring inspired CO₂ when CO₂ gas is being therapeutically administered³³

4.8 Graphic evaluation of the ventilator-patient interface; evaluation of the shape of the capnogram may be useful in detecting rebreathing of CO₂, obstructive pulmonary disease, waning

neuromuscular blockade ('curare cleft'), cardiogenic oscillations, esophageal intubation, cardiac arrest, and contamination of the monitor or sampling line with secretions or mucus.^{6,34}

4.9 Measurement of the volume of CO₂ elimination to assess metabolic rate and/or alveolar ventilation³⁵

CO2 MV 5.0 CONTRAINDICATIONS:

There are no absolute contraindications to capnography in mechanically ventilated patients provided that the data obtained are evaluated with consideration given to the patient's clinical condition.

CO2 MV 6.0 HAZARDS/COMPLICATIONS:

Capnography with a clinically approved device is a safe, noninvasive test, associated with few hazards. With mainstream analyzers, the use of too large a sampling window may introduce an excessive amount of dead space into the ventilator circuit.^{2,36} Care must be taken to minimize the amount of additional weight placed on the artificial airway by the addition of the sampling window or, in the case of a sidestream analyzer, the sampling line.

CO2 MV 7.0 LIMITATIONS OF PROCEDURE OR DEVICE:

Capnography, when performed using a device calibrated and operated as recommended by the manufacturer, has few limitations. It is important to note that although the capnograph provides valuable information about the efficiency of ventilation (as well as pulmonary, systemic, and coronary perfusion), it is not a replacement or substitute for assessing the P_aCO₂.^{20,34,37-41} The difference between P_{et}CO₂ and P_aCO₂ increases as dead-space volume increases. In fact, the difference between the P_aCO₂ and P_{et}CO₂ has been shown to vary within the same patient over time.⁴²⁻⁴⁷ Alterations in breathing pattern and tidal volume may introduce error into measurements designed to be made during stable, steady-state conditions.⁴⁸⁻⁵⁰ Interpretation of results must take into account the stability of physiologic parameters such as minute ventilation, tidal volume, cardiac output, ventilation/perfusion ratios and CO₂ body stores.^{51,52} Certain situations may affect the reliability of the capnogram. The extent to which

the reliability is affected varies somewhat among types of devices (infrared,⁵³ photoacoustic, mass spectrometry, and Raman spectrometry). Limitations include:

7.1 The composition of the respiratory gas mixture may affect the capnogram (depending on the measurement technology incorporated).

7.1.1 The infrared spectrum of CO₂ has some similarities to the spectra for both oxygen and nitrous oxide.⁵³ High concentrations of either or both oxygen or nitrous oxide may affect the capnogram, and, therefore, a correction factor should be incorporated into the calibration of any capnograph used in such a setting.⁵⁴

7.1.2 The reporting algorithm of some devices (primarily mass spectrometers) assumes that the only gases present in the sample are those that the device is capable of measuring.⁵⁵ When a gas that the mass spectrometer cannot detect (such as helium) is present, the reported values of CO₂ are incorrectly elevated in proportion to the concentration of helium present.

7.2 The breathing frequency may affect the capnograph. High breathing frequencies may exceed the response capabilities of the capnograph.⁵⁶ In addition, breathing frequency above 10 breaths/min has been shown to affect devices differently.⁵⁷

7.3 The presence of Freon (used as a propellant in metered dose inhalers) in the respiratory gas has been shown to artificially increase the CO₂ reading of mass spectrometers (ie, to show an apparent increase in [CO₂]). A similar effect has not yet been demonstrated with Raman or infrared spectrometers.⁵⁸

7.4 Contamination of the monitor or sampling system by secretions or condensate, a sample tube of excessive length, a sampling rate that is too high, or obstruction of the sampling chamber can lead to unreliable results.

7.5 Use of filters between the patient airway and the sampling line of the capnograph may lead to lowered P_{et}CO₂ readings.^{59,60}

7.6 Low cardiac output may cause a false negative result when attempting to verify endotracheal tube (ETT) position in the trachea. False positive results have been reported with ETT position in the pharynx and when antacids

and/or carbonated beverages are present in the stomach.^{18,61}

7.7 Decreased tidal volume delivery is possible during volume modes, some dual control modes, and time-cycled pressure limited ventilation with low continuous flowrates if the sampling flowrate of a sidestream analyzer is too high, especially in neonates and pediatrics.

7.8 Inaccurate measurement of expired CO₂ may be caused by leaks of gas from the patient/ventilator system preventing collection of expired gases, including

7.8.1 Leaks in the ventilator circuit⁶²

7.8.2 Leaks around tracheal tube cuffs or uncuffed tubes⁶²

CO2 MV 8.0 ASSESSMENT OF NEED:

Capnography is considered a standard of care during anesthesia.^{13,63} The American Society of Anesthesiologists has suggested that capnography be available for patients with acute ventilatory failure on mechanical ventilatory support.⁶³ The American College of Emergency Physicians recommends capnography as an adjunctive method to ensure proper ETT position.⁶⁴ The international guidelines for emergency cardiovascular care recommend use of capnography to verify ETT placement in all age groups.⁶⁵ Assessment of the need to use capnography with a specific patient should be guided by the clinical situation. The patient's primary cause of respiratory failure and the acuteness of his or her condition should be considered.

CO2 MV 9.0 ASSESSMENT OF OUTCOME:

Results should reflect the patient's condition and should validate the basis for ordering the monitoring. Documentation of results (along with all ventilatory and hemodynamic variables available), therapeutic interventions, and/or clinical decisions made based on the capnogram should be included in the patient's chart.

CO2 MV 10.0 RESOURCES:

10.1 Equipment: The capnograph and accessories (eg, airway adapter, sampling tube, depending on capnograph). The capnograph should be calibrated as recommended by the manufacturer.

10.2 Personnel: Licensed or credentialed respiratory care practitioners or individuals with similar credentials (eg, MD, RN) who have the necessary training and demonstrated skills to correctly calibrate and evaluate the capnograph, assess the patient and the patient-ventilator system, and the ability to exercise appropriate clinical judgment.

CO2 MV 11.0 MONITORING:

During capnography the following should be considered and monitored:

11.1 Ventilatory variables: tidal volume, respiratory rate, positive end-expiratory pressure, inspiratory-to-expiratory time ratio (I:E), peak airway pressure, and concentrations of respiratory gas mixture^{36,37}

11.2 Hemodynamic variables: systemic and pulmonary blood pressures, cardiac output, shunt, and ventilation-perfusion imbalances^{36,37}

CO2 MV 12.0 FREQUENCY:

Capnography (or, at least, capnometry) should be available during endotracheal intubation.^{13,14,16,37,64} Capnography is not indicated for every mechanically ventilated patient; however, when it is used, the measurement period should be long enough to allow determination of the P_aCO₂-P_{et}CO₂ difference, note changes in P_aCO₂-P_{et}CO₂ difference as a result of therapy, and allow interpretation of observed trends.

CO2 MV 13.0 INFECTION CONTROL:

No specific precautions are necessary although Standard Precautions (as described by the Centers for Disease Control & Prevention)⁶⁶ and precautions designed to limit the spread of tuberculosis⁶⁷ should always be implemented during patient care.

13.1 The sensor (the portion of the device contacting the patient's airway) should be subjected to high-level disinfection between patients, according to the manufacturer's recommendations.

13.2 The monitor (the portion not contacting the patient or the patient's airway) should be cleaned as needed according to manufacturer's recommendations.

*Revised by Charles D McArthur RRT RPFT, Im -
manuel St Joseph's - Mayo Health System, Manka -
to, Minnesota. and approved by the 2003 CPG
Steering Committee*

Original Publication: Respir Care 1995;40(12):1321-1324.

REFERENCES

- Hess DR, Branson RD. Noninvasive respiratory monitoring equipment. In: Branson RD, Hess DR, Chatburn RL, editors. Respiratory care equipment. Philadelphia: Lippincott; 1994:184-216.
- Block FE Jr, McDonald JS. Sidestream versus mainstream carbon dioxide analyzers. *J Clin Monit* 1992;8(2):139-141.
- O'Flaherty D. Capnometry. London: BMJ Publishing Group; 1994:21-54.
- VanWagenen RA, Westenskow DR, Benner RE, Gregonis DE, Coleman DL. Dedicated monitoring of anesthetic and respiratory gases by Raman scattering. *J Clin Monit* 1986;2(4):215-222.
- Cinnella G, Dambrosio M, Brienza N, Giuliani R, Bruno F, Fiore T, Brienza A. Independent lung ventilation in patients with unilateral pulmonary contusion: monitoring with compliance and etCO_2 . *Intensive Care Med* 2001;27(12):1860-1867.
- Carlson GC, Ray C Jr, Miodownik S, Kopec I, Groeger JS. Capnography in mechanically ventilated patients. *Crit Care Med* 1988;16(5):550-556.
- Gravenstein N, Good ML. Noninvasive assessment of cardiopulmonary function. In: Civetta JM, Taylor RW, Kirby RR, editors. Critical care. Philadelphia: Lippincott; 1988:291-311.
- Yamanaka MK, Sue DY. Comparison of arterial-end-tidal P_{CO_2} difference and dead space/tidal volume ratio in respiratory failure. *Chest* 1987;92(5):832-835.
- Poppius H, Korhonen O, Viljanen AA, Kreuz KE. Arterial to end-tidal CO_2 difference in respiratory disease. *Scand J Respir Dis* 1975;56(5):254-262.
- Burrows FA. Physiologic dead space, venous admixture, and the arterial to end-tidal carbon dioxide difference in infants and children undergoing cardiac surgery. *Anesthesiology* 1989;70(2):219-225.
- Kafer ER. Respiratory function in pulmonary thromboembolic disease. *Am J Med* 1969;47(6):904-915.
- Sanders AB, Atlas M, Ewy GA, Kern KB, Bragg S. Expired P_{CO_2} as an index of coronary perfusion pressure. *Am J Emerg Med* 1985;3(2):147-149.
- Eichhorn JH, Cooper JB, Cullen DJ, Maier WR, Philip JH, Seeman RG. Standards for patient monitoring during anesthesia at Harvard Medical School. *JAMA* 1986;256(8):1017-1020.
- Birmingham PK, Cheney FW, Ward RJ. Esophageal intubation: a review of detection techniques. *Anesth Analg* 1986;65(8):886-891.
- Sanders AB. Capnometry in emergency medicine. *Ann Emerg Med* 1989;18(12):1287-1290.
- Murray IP, Modell JH. Early detection of endotracheal tube accidents by monitoring carbon dioxide concentration in respiratory gas. *Anesthesiology* 1983;59(4):344-346.
- Roberts WA, Maniscalco WM, Cohen AR, Litman RS, Chhibber A. The use of capnography for recognition of esophageal intubation in the neonatal intensive care unit. *Pediatr Pulmonol* 1995;19(5):262-268.
- Li J. Capnography alone is imperfect for endotracheal tube placement confirmation during emergency intubation. *J Emerg Med* 2001;20(3):223-229.
- Goldberg JS, Rawle PR, Zehnder JL, Sladen RN. Colorimetric end-tidal carbon dioxide monitoring for tracheal intubation. *Anesth Analg* 1990;70(2):191-194.
- Graybeal JM, Russell GB. Capnometry in the surgical ICU: an analysis of the arterial-to-end-tidal carbon dioxide difference. *Respir Care* 1993;38(8):923-928.
- Braman SS, Dunn SM, Amico CA, Millman RP. Complications of intrahospital transport of critically ill patients. *Ann Internal Med* 1987;107(4):469-473.
- Kalenda Z. The capnogram as a guide to the efficacy of cardiac massage. *Resuscitation* 1978;6(4):259-263.
- Sanders AB, Ewy GA, Bragg S, Atlas M, Kern KB. Expired P_{CO_2} as prognostic indicator of successful resuscitation from cardiac arrest. *Ann Emerg Med* 1985;14(10):948-952.
- Shibutani K, Muraoka M, Shirasaki S, Kubal K, Sanchala VT, Gupte P. Do changes in end-tidal P_{CO_2} quantitatively reflect changes in cardiac output? *Anesth Analg* 1994;79(5):829-833.
- Weil MH, Bisera J, Trevino RP, Rackow EC. Cardiac output and end-tidal carbon dioxide. *Crit Care Med* 1985;13(11):907-909.
- Levine RL, Wayne MA, Miller CC. End-tidal carbon dioxide and outcome of out-of-hospital cardiac arrest. *N Engl J Med* 1997;337(5):301-306.
- de Abreu MG, Quintel M, Ragaller M, Albrecht DM. Partial carbon dioxide rebreathing: a reliable technique for noninvasive measurement of nonshunted pulmonary capillary blood flow. *Crit Care Med* 1997;25(4):675-683.
- van Heerden PV, Baker S, Lim SI, Weidman C, Bulsara M. Clinical evaluation of the non-invasive cardiac output (NICO) monitor in the intensive care unit. *Anaesth Intensive Care* 2000;28(4):427-430.
- de Abreu MG, Geiger S, Winkler T, Ragaller M, Pfeiffer T, Leutheuser D, Albrecht DM. Evaluation of a new device for noninvasive measurement of nonshunted pulmonary capillary blood flow in patients with acute lung injury. *Intensive Care Med* 2002;28(3):318-323.
- Rodger MA, Jones G, Rasuli P, Raymond F, Djunaedi H, Bredeson CN, Wells PS. Steady-state end-tidal alveolar dead space fraction and D-dimer: bedside tests to exclude pulmonary embolism. *Chest* 2001;120(1):115-119.
- Kline JA, Johns KL, Colucciello SA, Israel EG. New diagnostic tests for pulmonary embolism. *Ann Emerg Med* 2000;35(2):168-180.
- Colman Y, Krauss B. Microstream capnography technol-

- ogy: a new approach to an old problem. *J Clin Monit Comput* 1999;15(6):403-409.
33. Fatigante L, Cartei F, Ducci F, Marini C, Prediletto R, Caciagli P, Laddaga M. Carbogen breathing in patients with glioblastoma multiforme submitted to radiotherapy: assessment of gas exchange parameters. *Acta Oncol* 1994;33(7):807-811.
 34. Bhavani-Shankar K, Moseley H, Kumar AY, Delphi Y. Capnometry and anesthesia. *Can J Anesth* 1992;39(6):617-632.
 35. Taskar V, John J, Larsson A, Wetterberg T, Jonson B. Dynamics of carbon dioxide elimination following ventilator resetting. *Chest* 1995;108(1):196-202.
 36. Szaflarski NL, Cohen NH. Use of capnography in critically ill adults. *Heart Lung* 1991;20(4):363-374. *Erratum* 1991;20(6):630
 37. Hess D. Capnometry and capnography: technical aspects, physiologic aspects, and clinical applications. *Respir Care* 1990;35(6):557-573.
 38. Isert P. Control of carbon dioxide levels during neuroanesthesia: current practice and an appraisal of our reliance upon capnography. *Anaesth Intensive Care* 1994;22(4):435-441.
 39. Jellinek H, Hiesmayr M, Simon P, Klepetko W, Haider W. Arterial to end-tidal CO₂ tension difference after bilateral lung transplantation. *Crit Care Med* 1993;21(7):1035-1040.
 40. Laffon M, Gouchet A, Sitbon P, Guicheteau V, Biyick E, Duchalais A, Mercier C. Difference between arterial and end-tidal carbon dioxide pressures during laparoscopy in paediatric patients. *Can J Anaesth* 1998;45(6):561-563.
 41. Kerr ME, Zempsky J, Sereika S, Orndoff P, Rudy EB. Relationship between arterial carbon dioxide and end-tidal carbon dioxide in mechanically ventilated adults with severe head trauma. *Crit Care Med* 1996;24(5):785-790.
 42. Russell GB, Graybeal JM. End-tidal carbon dioxide as an indicator of arterial carbon dioxide in neurointensive care patients. *J Neurosurg Anesth* 1992;4:245-249.
 43. Russell GB, Graybeal JM. Reliability of the arterial to end-tidal carbon dioxide gradient in mechanically ventilated patients with multisystem trauma. *J Trauma* 1994;36(3):317-322.
 44. Russell GB, Graybeal JM. The arterial to end-tidal carbon dioxide difference in neurosurgical patients during craniotomy. *Anesth Analg* 1995;81(4):806-810.
 45. Hess D, Schlottag A, Levin B, Mathai J, Rexrode WO. An evaluation of the usefulness of end-tidal P_{CO₂} to aid weaning from mechanical ventilation following cardiac surgery. *Respir Care* 1991;36(8):837-843.
 46. Grenier B, Verchere E, Mesli A, Dubreuil M, Siao D, Vandendriessche M, et al. Capnography monitoring during neurosurgery: reliability in relation to various intraoperative positions. *Anesth Analg* 1999;88(1):43-48.
 47. Seguin P, Bleichner JP, Branger B, Guillou YM, Feuillu A, Malledant Y. The measurement of end-tidal carbon dioxide (P_{etCO₂}) is not a significant parameter to monitor in patients with severe traumatic brain injury. *Can J Anaesth* 2001;48(4):396-400.
 48. Wessel HU, Stout RL, Bastanier CK, Paul MH. Breath-by-breath variation of FRC: effect on V_{O₂} and V_{CO₂} measured at the mouth. *J Appl Physiol* 1979;46(6):1122-1126.
 49. Gamma de Abreu M, Melo MF, Giannella-Neto A. Pulmonary capillary blood flow by partial CO₂ rebreathing: importance of the regularity of the respiratory pattern. *Clin Physiol* 2000;20(5):388-398.
 50. Brandi LS, Bertolini R, Santini L, Cavani S. Effects of ventilator resetting on indirect calorimetry measurement in the critically ill surgical patient. *Crit Care Med* 1999;27(3):531-539.
 51. Farhi LE, Rahn H. Dynamics change in carbon dioxide stores. *Anesthesiology* 1960;21:604-614.
 52. Damask MC, Askanazi J, Weissman C, Elwyn DH, Kinney JM. Artifacts in measurement of resting energy expenditure. *Crit Care Med* 1983;11(9):750-752.
 53. Ammann ECB, Galvin RD. Problems associated with the determination of carbon dioxide by infrared absorption. *J Appl Physiol* 1968;25(3):333-335.
 54. Severinghaus JW, Larson CP, Eger EI. Correction factors for infrared carbon dioxide pressure broadening by nitrogen, nitrous oxide and cyclopropane. *Anesthesiology* 1961;22:429-432.
 55. Perkin-Elmer MGA 1100 Operation and maintenance manual. 1982:6-26.
 56. From RP, Scamman FL. Ventilatory frequency influences accuracy of end-tidal CO₂ measurements: analysis of seven capnometers. *Anesth Analg* 1988;67(9):884-886.
 57. Graybeal JM, Russell GB. Relative agreement between Raman and mass spectrometry for measuring end-tidal carbon dioxide. *Respir Care* 1994;39(3):190-194.
 58. Elliott WR, Raemer DB, Goldman DB, Philip JH. The effects of bronchodilator-inhaler aerosol propellants on respiratory gas monitors. *J Clin Monit* 1991;7(2):175-180.
 59. Hardman JG, Curran J, Mahajan RP. End-tidal carbon dioxide measurement and breathing system filters. *Anaesthesia* 1997;52(7):646-648.
 60. Hardman JG, Mahajan RP, Curran J. The influence of breathing system filters on paediatric capnography. *Paediatr Anaesth* 1999;9(1):35-38.
 61. Sum Ping ST, Mehta MP, Symreng T. Reliability of capnography in identifying esophageal intubation with carbonated beverage or antacid in the stomach. *Anesth Analg* 1991;73(3):333-337.
 62. Branson RD. The measurement of energy expenditure: instrumentation, practical considerations, and clinical application. *Respir Care* 1990;35(7):640-656; discussion 656-659.
 63. American Society of Anesthesiologists: Standards for basic anesthetic monitoring. 1998. Available at: <http://www.asahq.org/>
 64. American College of Emergency Physicians. Clinical Policies Committee. Verification of endotracheal tube placement. *Ann Emerg Med* 2002;40(5):551-552.

65. Guidelines 2000 for Cardiopulmonary resuscitation and emergency cardiovascular care: international consensus on science. *Circulation* 2000;102(8):I1-I370.
66. Bolyard EA, Tablan OC, Williams WW, Pearson ML, Shapiro CN, Deitchmann SD. Guideline for infection control in healthcare personnel, 1998. Hospital Infection Control Practices Advisory Committee. *Infect Control Hosp Epidemiol* 1998;19(6):4407-4462. *Erratum in: Infect Control Hosp Epidemiol* 1998;19(7):4493
67. Guidelines for preventing the transmission of *Mycobacterium tuberculosis* in health-care facilities, 1994. Centers for Disease Control and Prevention. *MMWR Recomm Rep* 1994;43(RR-13):1-132.

Interested persons may photocopy these Guidelines for noncommercial purposes of scientific or educational advancement. Please credit AARC and RESPIRATORY CARE Journal.

All of the AARC CPGs may be downloaded at no charge from
<http://www.rcjournal.com/online/>