



Supplemental digital content Figure 1: Pressure response for systems during leakage. Pressure change during constant leakage (1-4 l/min) applied between breath number 30 and 45 (blue bar) at a set CPAP of 4 cm H₂O. Amplitude of pressure swings represents highest and lowest pressure for each breath. Delivered mean CPAP indicated by black diamonds.